

Analytical Results	
<b>% Oxygen</b> Mean = 0.0033 St Dev = 0.0004 Exp Uncertainty = 0.0008 k=2, @95% CI, n=61	<b>% Nitrogen</b> Mean = 0.0222 St Dev = 0.0011 Exp Uncertainty = 0.0025 k=2, @95% CI, n=61
Primary (NMI) Reference Standards Employed: NIST SRM – 1099, 1096, 1098, 50c, 1095 JSS – 605-11, 610-10, 611-11 EURO – 231-2, 026-1	
<b>Method of Analysis: ASTM E1019-18</b>	

*\*The analytical results above are provided by an accredited reference material manufacturer with a current certification in ISO 17025 and 17034.*

The intended use of this Reference Material (RM) is for the calibration and validation of inert gas fusion analyzers with infrared (oxygen) and thermal conductivity (nitrogen) detection as described in the above ASTM methods.

The minimum sample size to perform this intended use is 1 pin (1.0g nominal). Multiple pins may be used as required. Refer to test method recommendations for an appropriate sample size.

The Period of Validity for this RM is not able to be determined and should be reviewed every 20 years after the date below.

This bottle contains 100g of 1.0g (nominal) pins to be used per the test method you follow. Keep sealed tightly and store under normal laboratory conditions.

Refer to your test methods and or manufacturer manual for expanded uncertainties, repeatability/reproducibility factors.

For good laboratory practice, we recommend that all reference materials be verified as fit for purpose prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd. be liable for incidental or consequential damages.

**Certified on the 10<sup>th</sup> of May 2024.**

Elemental Microanalysis Ltd