

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

Certificate of Analysis**Part No. B2369****Residual Oil Standard**

Certificate Number	814614
Date	29 September 2014
Page	1 of 1

Weight Percent Sulphur = 0.99
Standard Deviation = 0.04
(k=2, 95% confidence limit)

Methods used for analysis: ASTM D-4294-10, ARI-078

NMI Standard(s) used: NIST 1622d, 1619a

This standard was produced gravimetrically using high purity materials, with balances calibrated and checked by precision NIST traceable weights. The above ASTM method and NMI references listed were used for testing. The uncertainty value represents the 95% confidence limit (k=2) derived from analysis (n=40), and possible errors in weighing. Normal ASTM procedures should be employed when using this standard. This includes the reproducibility and repeatability uncertainty for the method you wish to employ. The material used in production of this standard was sampled in accordance with ARI 032. The samples used for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard.

Notes

Before use, the contents of the bottle should be mixed by vigorous shaking. Any exposure to air and light should be kept to a minimum. Keep sealed and store upright under normal laboratory conditions. This bottle consists of 100ml to be used as per the test method. Unopened and stored properly, this product has an indefinite shelf life. Once opened, this certificate is valid for two years.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Material (working standard), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

This CRM is valid for two years from the date of opening

Elemental Microanalysis Limited

July 7, 2014