

Certificate of Analysis

(Isotopic Composition)

High Organic Sediment Standard OAS
Cat No. B2151 – Certificate No. 295716

General

This Organic Analytical Standard (OAS) consists of a homogenous batch of High Organic Content Sediment standard for use as a routine working microanalytical standard for the determination of the Carbon and Nitrogen isotopes ^{13}C , ^{15}N .

Certified Values and Uncertainty

The uncertainty in the certified value is expressed as expanded uncertainty, U, at 95% confidence and is calculated in accordance with ISO/IEC17025 according to GUM (Guidelines to Uncertainty in Measurement). Confidence limits include those due to sampling variation, weighing, calibration and measurement errors. The certified values are based upon the results of 5 determinations

The certified value for Carbon was determined by EA-IRMS and traceable to IAEA-CH-6 (IAEA, Vienna)

The certified value for Nitrogen was determined by EA-IRMS and traceable to IAEA-N-1 (IAEA, Vienna)

<i>Isotope</i>	<i>Certified value</i> (‰)	<i>Uncertainty</i> (+/- ‰)
$\delta^{13}\text{C}_{\text{V-PDB}}$	-28.85	0.10
$\delta^{15}\text{N}_{\text{AIR}}$	+4.32	0.20

Expiration of Certification

The certification of this OAS is valid until 1 May 2021 within the measurement uncertainties specified

Storage and use

This OAS should be stored between 20°C to 25°C and should be kept tightly sealed away from light and moisture. It is non-hygroscopic under normal conditions and can be used without preliminary drying.

Certification Information

The technical aspects involved in the preparation, certification and issuance of this (In)Organic Analytical Standard (IAS/OAS) were carried out at Elemental Microanalysis Ltd, Okehampton, Devon EX20 1UB UK, Tel +44 1837 54446, Fax +44 1837 54544, Email enquires@microanalysis.co.uk

For and on behalf of
Elemental Microanalysis Ltd

Jon Davies
Technical Director



Certificate of Analysis

(Elemental Composition)

High Organic Sediment Standard OAS
Cat No. B2151 – Certificate No. 295716

General

This Organic Analytical Standard (OAS) consists of a homogenous batch of High Organic Content Sediment standard for use as a routine working microanalytical standard for the determination of Carbon, Nitrogen and Sulphur.

Certified Values and Uncertainty

The uncertainty in the certified value is expressed as expanded uncertainty, U, at 95% confidence and is calculated in accordance with ISO/IEC17025 according to GUM (Guidelines to Uncertainty in Measurement). Confidence limits include those due to sampling variation, weighing, calibration and measurement errors. The certified values are based upon the results of 30 determinations.

The certified values for Carbon and Nitrogen were determined by elemental analyser calibrated to Cystine 143d from National Institute of Standards and Technology (NIST), Maryland, USA.

The certified value for Sulphur was determined by elemental analyser calibrated to Cystine 143d from National Institute of Standards and Technology (NIST), Maryland, USA.

<i>Element</i>	<i>Certified value</i> (% w/w)	<i>Uncertainty</i> (+/- %)
Carbon	7.45	0.14
Nitrogen	0.52	0.02
Sulphur	0.62	0.02

Expiration of Certification

The certification of this OAS is valid until 1 May 2021 within the measurement uncertainties specified

Storage and use

This OAS should be stored between 20°C to 25°C and should be kept tightly sealed away from light and moisture. It is non-hygroscopic under normal conditions and can be used without preliminary drying.

Certification Information

The technical aspects involved in the preparation, certification and issuance of this (In)Organic Analytical Standard (IAS/OAS) were carried out at Elemental Microanalysis Ltd, Okehampton, Devon EX20 1UB UK, Tel +44 1837 54446, Fax +44 1837 54544, Email info@microanalysis.co.uk.

For and on behalf of
Elemental Microanalysis Ltd

Jon Davies
Technical Director